



PubMed	Nucleotide	Protein	Genome	Structure	PMC	Taxonomy	OMIM
Search PubMed	<input type="text" value="for"/>						Go
Limits Preview/Index History Clipboard Details							
Display	Abstract	<input type="text" value="Show: 20"/>	<input type="text" value="Sort"/>	<input type="text" value="Send to: Text"/>			

☐ 1: J Biomol NMR. 2002 Sep;24(1):41-50.

[Related Articles, Links](#)

Entrez PubMed

Novel projected 4D triple resonance experiments for polypeptide backbone chemical shift assignment.

Xia Y, Arrowsmith CH, Szyperski T.

PubMed Services

Ontario Cancer Institute and Department of Medical Biophysics, The University of Toronto, Canada.

Related
Resources

Here we present a novel suite of projected 4D triple-resonance NMR experiments for efficient sequential assignment of polypeptide backbone chemical shifts in $^{13}\text{C}/^{15}\text{N}$ doubly labeled proteins. In the 3D HNN[CAHA] and 3D HNN(CO)[CAHA] experiments, the $^{13}\text{C}(\alpha)$ and $^1\text{H}(\alpha)$ chemical shifts evolve in a common dimension and are simultaneously detected in quadrature. These experiments are particularly useful for the assignment of glycine-rich polypeptide segments. Appropriate setting of the ^1H radiofrequency carrier allows one to place cross peaks correlating either backbone $^{15}\text{N}/^1\text{H}(\text{N})/^{13}\text{C}(\alpha)$ or $^{15}\text{N}/^1\text{H}(\text{N})/^1\text{H}(\alpha)$ chemical shifts in separate spectral regions. Hence, peak overlap is not increased when compared with the conventional 3D HNNCA and HNN(CA)HA. 3D HNN[CAHA] and 3D HNN(CO)[CAHA] are complemented by 3D reduced-dimensionality (RD) HNN COCA and NNCACO, where $^{13}\text{C}(\alpha)$ and $^{13}\text{C}'$ chemical shifts evolve in a common dimension. The $^{13}\text{C}(\alpha)$ shift is detected in quadrature, which yields peak pairs encoding the $^{13}\text{C}'$ chemical shift in an in-phase splitting. This suite of four experiments promises to be of value for automated high-throughput NMR structure determination in structural genomics, where the requirement to independently sample many indirect dimensions in a large number of NMR experiments may prevent one from accurately adjusting NMR measurement times to spectrometer sensitivity.

PMID: 12449417 [PubMed - indexed for MEDLINE]

Display	Abstract	<input type="text" value="Show: 20"/>	<input type="text" value="Sort"/>	<input type="text" value="Send to: Text"/>
---------	----------	---------------------------------------	-----------------------------------	--